

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060519\
 Data File : VX010097.D
 Acq On : 06 Jun 2019 00:15
 Operator : JC/SP
 Sample : K3203-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1-TB08-190604

Quant Time: Jun 06 07:40:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	172886	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	275408	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	249332	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113248	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	112082	45.89	ug/l	0.00
Spiked Amount			Recovery	=	91.78%	
35) Dibromofluoromethane	5.50	113	83144	48.02	ug/l	0.00
Spiked Amount			Recovery	=	96.04%	
50) Toluene-d8	8.71	98	346699	49.31	ug/l	0.00
Spiked Amount			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	11.13	95	118990	47.12	ug/l	0.00
Spiked Amount			Recovery	=	94.24%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060519\
Data File : VX010097.D
Acq On : 06 Jun 2019 00:15
Operator : JC/SP
Sample : K3203-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1-TB08-190604

Quant Time: Jun 06 07:40:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

