


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX060519\
 Data File : VX010064.D
 Acq On : 05 Jun 2019 09:08
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Title : SW846 8260
 Last Update : Wed May 29 04:00:56 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.8	7800	PASS
75	95	30	60	51.8	19432	PASS
95	95	100	100	100.0	37522	PASS
96	95	5	9	6.5	2446	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	82.2	30856	PASS
175	174	5	9	7.8	2400	PASS
176	174	95	101	96.4	29757	PASS
177	176	5	9	7.4	2188	PASS