

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060520\
 Data File : VX016608.D
 Acq On : 05 Jun 2020 14:01
 Operator : JC/SP
 Sample : L2894-11DL 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW033-200601DL

Quant Time: Jun 06 07:29:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	231392	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	370041	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	372042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	188455	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	139170	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
35) Dibromofluoromethane	5.46	113	112982	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
50) Toluene-d8	8.70	98	456229	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	11.12	95	183023	53.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.82%	
Target Compounds						
4) Vinyl Chloride	1.39	62	79014	27.151	ug/l	98
27) cis-1,2-Dichloroethene	4.57	96	145814	51.179	ug/l	89
65) Chlorobenzene	10.12	112	4402	0.580	ug/l	97
91) 1,2-Dichlorobenzene	12.38	146	1231	0.214	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060520\
Data File : VX016608.D
Acq On : 05 Jun 2020 14:01
Operator : JC/SP
Sample : L2894-11DL 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW033-200601DL

Quant Time: Jun 06 07:29:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

