

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060520\
 Data File : VX016611.D
 Acq On : 05 Jun 2020 15:09
 Operator : JC/SP
 Sample : L2894-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012RW043-200602

Quant Time: Jun 06 06:38:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	227492	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	370152	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	371429	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	190037	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	138365	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
35) Dibromofluoromethane	5.47	113	113701	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
50) Toluene-d8	8.70	98	454125	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
62) 4-Bromofluorobenzene	11.12	95	181170	52.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.70%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	2692	0.961	ug/l	89
65) Chlorobenzene	10.12	112	3593	0.474	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060520\
Data File : VX016611.D
Acq On : 05 Jun 2020 15:09
Operator : JC/SP
Sample : L2894-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012RW043-200602

Quant Time: Jun 06 06:38:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

