

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060520\
 Data File : VX016615.D
 Acq On : 05 Jun 2020 16:40
 Operator : JC/SP
 Sample : L2894-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW026-200601

Quant Time: Jun 06 06:56:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	221962	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	351756	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	361278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	185771	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	133695	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
35) Dibromofluoromethane	5.46	113	107933	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
50) Toluene-d8	8.70	98	436683	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
62) 4-Bromofluorobenzene	11.12	95	175584	53.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.80%	
Target Compounds						
16) Acetone	2.42	43	4654	3.854	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	1197	0.438	ug/l	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060520\
Data File : VX016615.D
Acq On : 05 Jun 2020 16:40
Operator : JC/SP
Sample : L2894-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW026-200601

Quant Time: Jun 06 06:56:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

