

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060520\
 Data File : VX016618.D
 Acq On : 05 Jun 2020 17:48
 Operator : JC/SP
 Sample : L2894-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW147-200602

Quant Time: Jun 06 07:05:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	225586	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	356795	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	368396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	194340	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	135182	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
35) Dibromofluoromethane	5.47	113	109563	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
50) Toluene-d8	8.70	98	447991	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
62) 4-Bromofluorobenzene	11.12	95	182207	55.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.30%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.39	62	2514	0.886	ug/l	100
16) Acetone	2.42	43	2510	2.045	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	19635	7.069	ug/l	87
44) Trichloroethene	7.20	130	1089	0.341	ug/l	92
65) Chlorobenzene	10.12	112	7713	1.027	ug/l	96
91) 1,2-Dichlorobenzene	12.38	146	2898	0.489	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060520\
Data File : VX016618.D
Acq On : 05 Jun 2020 17:48
Operator : JC/SP
Sample : L2894-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW147-200602

Quant Time: Jun 06 07:05:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

