

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060520\
 Data File : VX016622.D
 Acq On : 05 Jun 2020 19:19
 Operator : JC/SP
 Sample : L2926-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TANK-1-20200604

Quant Time: Jun 06 07:14:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	213068	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	341523	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	352405	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	176623	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	129452	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
35) Dibromofluoromethane	5.47	113	104856	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	8.70	98	425216	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
62) 4-Bromofluorobenzene	11.12	95	171507	54.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.46%	
Target Compounds						
16) Acetone	2.42	43	2625	2.264	ug/l	95
27) cis-1,2-Dichloroethene	4.57	96	3351	1.277	ug/l	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060520\
Data File : VX016622.D
Acq On : 05 Jun 2020 19:19
Operator : JC/SP
Sample : L2926-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TANK-1-20200604

Quant Time: Jun 06 07:14:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

