

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
 Data File : VX035994.D
 Acq On : 05 Jun 2023 10:49
 Operator : JC/MD
 Sample : VX0605WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0605WBL01

Quant Time: Jun 06 06:41:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	238462	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	386439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	352545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	165321	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	175867	45.417	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.840%
35) Dibromofluoromethane	5.385	113	120500	46.361	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.720%
50) Toluene-d8	8.647	98	457618	49.057	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.120%
62) 4-Bromofluorobenzene	11.079	95	177636	48.239	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
Data File : VX035994.D
Acq On : 05 Jun 2023 10:49
Operator : JC/MD
Sample : VX0605WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0605WBL01

Quant Time: Jun 06 06:41:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

