

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
 Data File : VX035998.D
 Acq On : 05 Jun 2023 12:58
 Operator : JC/MD
 Sample : 03022-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-TB-20230525

Quant Time: Jun 06 06:43:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	230524	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	386268	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	341407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	155866	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	166523	44.485	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.960%
35) Dibromofluoromethane	5.385	113	117829	45.353	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.700%
50) Toluene-d8	8.647	98	450722	48.339	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.680%
62) 4-Bromofluorobenzene	11.079	95	171933	46.711	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.420%
Target Compounds						
16) Acetone	2.422	43	1707	1.105	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
Data File : VX035998.D
Acq On : 05 Jun 2023 12:58
Operator : JC/MD
Sample : 03022-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-TB-20230525

Quant Time: Jun 06 06:43:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

