

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060523\  
 Data File : VX036000.D  
 Acq On : 05 Jun 2023 13:45  
 Operator : JC/MD  
 Sample : 03022-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BP-VPB-RW11-GW-518-520

Quant Time: Jun 06 06:44:21 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X053123W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 31 12:50:30 2023  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.550	168	237559	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	380239	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	338946	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	155828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	160514	41.610	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.220%
35) Dibromofluoromethane	5.385	113	113791	44.494	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.980%
50) Toluene-d8	8.647	98	447730	48.780	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.560%
62) 4-Bromofluorobenzene	11.079	95	169689	46.833	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.660%
Target Compounds						
16) Acetone	2.392	43	2519	1.583	ug/l	95
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060523\  
 Data File : VX036000.D  
 Acq On : 05 Jun 2023 13:45  
 Operator : JC/MD  
 Sample : 03022-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_X  
**ClientSampleId :**  
 BP-VPB-RW11-GW-518-520

Quant Time: Jun 06 06:44:21 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X053123W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 31 12:50:30 2023  
 Response via : Initial Calibration

