

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
 Data File : VX036003.D
 Acq On : 05 Jun 2023 14:55
 Operator : JC/MD
 Sample : 03022-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-DUP-20230531

Quant Time: Jun 06 06:45:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	228700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	377642	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	339005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	148889	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	166374	44.799	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.600%
35) Dibromofluoromethane	5.385	113	113874	44.832	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.660%
50) Toluene-d8	8.647	98	444467	48.757	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.520%
62) 4-Bromofluorobenzene	11.079	95	172477	47.929	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.860%
Target Compounds						
16) Acetone	2.404	43	6558	4.280	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
Data File : VX036003.D
Acq On : 05 Jun 2023 14:55
Operator : JC/MD
Sample : 03022-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-DUP-20230531

Quant Time: Jun 06 06:45:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

