

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
 Data File : VX036004.D
 Acq On : 05 Jun 2023 15:19
 Operator : JC/MD
 Sample : 03022-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-GW-598-600

Quant Time: Jun 06 06:45:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	243512	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	394303	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	348567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162488	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	166040	41.990	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.980%
35) Dibromofluoromethane	5.385	113	119870	45.199	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.400%
50) Toluene-d8	8.647	98	462260	48.567	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.140%
62) 4-Bromofluorobenzene	11.079	95	175104	46.603	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.200%
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	7704	4.722	ug/l	99
20) Methylene Chloride	2.788	84	772	0.246	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
Data File : VX036004.D
Acq On : 05 Jun 2023 15:19
Operator : JC/MD
Sample : 03022-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-GW-598-600

Quant Time: Jun 06 06:45:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

