

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
 Data File : VX036005.D
 Acq On : 05 Jun 2023 15:42
 Operator : JC/MD
 Sample : 03022-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-GW-618-620

Quant Time: Jun 06 06:45:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	237914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	384043	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	328745	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151727	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	159104	41.183	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.360%
35) Dibromofluoromethane	5.379	113	116110	44.951	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.900%
50) Toluene-d8	8.647	98	431385	46.534	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.060%
62) 4-Bromofluorobenzene	11.079	95	164514	44.955	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.900%
Target Compounds						
16) Acetone	2.404	43	7286	4.571	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
Data File : VX036005.D
Acq On : 05 Jun 2023 15:42
Operator : JC/MD
Sample : 03022-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-GW-618-620

Quant Time: Jun 06 06:45:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

