

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
 Data File : VX036013.D
 Acq On : 05 Jun 2023 19:13
 Operator : JC/MD
 Sample : 03027-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 06 06:48:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	224827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	358381	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	311536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136152	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	153467	42.036	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.080%
35) Dibromofluoromethane	5.385	113	109583	45.462	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.920%
50) Toluene-d8	8.647	98	420216	48.575	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.140%
62) 4-Bromofluorobenzene	11.079	95	155255	45.462	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.920%
Target Compounds						
					Qvalue	
16) Acetone	2.416	43	1202	0.798	ug/l #	78
20) Methylene Chloride	2.788	84	1282	0.442	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060523\
Data File : VX036013.D
Acq On : 05 Jun 2023 19:13
Operator : JC/MD
Sample : 03027-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 06 06:48:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

