

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022482.D
 Acq On : 06 Jun 2021 16:59
 Operator : JC/MD
 Sample : PB136837TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136837TB

Quant Time: Jun 07 04:25:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	132573	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39506	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53374	52.514	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.020%
35) Dibromofluoromethane	5.397	113	39894	47.452	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.900%
50) Toluene-d8	8.653	98	159303	49.190	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.380%
62) 4-Bromofluorobenzene	11.085	95	51305	42.860	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.720%
Target Compounds						
16) Acetone	2.392	43	3666	8.284	ug/l #	87
20) Methylene Chloride	2.794	84	2787	2.836	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022482.D
Acq On : 06 Jun 2021 16:59
Operator : JC/MD
Sample : PB136837TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136837TB

Quant Time: Jun 07 04:25:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

