

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022484.D
 Acq On : 06 Jun 2021 17:46
 Operator : JC/MD
 Sample : PB136837ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136837ZHE#12

Quant Time: Jun 07 04:26:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64857	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	129360	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41001	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53176	53.151	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.300%			
35) Dibromofluoromethane	5.397	113	38513	46.948	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.900%			
50) Toluene-d8	8.653	98	160143	50.677	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.360%			
62) 4-Bromofluorobenzene	11.085	95	53687	45.964	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 91.920%			
Target Compounds						
16) Acetone	2.386	43	3837	8.808	ug/l #	81
20) Methylene Chloride	2.794	84	2753	2.846	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022484.D
Acq On : 06 Jun 2021 17:46
Operator : JC/MD
Sample : PB136837ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136837ZHE#12

Quant Time: Jun 07 04:26:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

