

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022485.D
 Acq On : 06 Jun 2021 18:09
 Operator : JC/MD
 Sample : PB136837ZHE#13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136837ZHE#13

Quant Time: Jun 07 04:26:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65416	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	132661	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	117904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53529	53.047	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.100%
35) Dibromofluoromethane	5.391	113	38905	46.245	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.500%
50) Toluene-d8	8.653	98	160436	49.507	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.020%
62) 4-Bromofluorobenzene	11.085	95	56574	47.231	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.460%
Target Compounds						
16) Acetone	2.386	43	3956	9.003	ug/l #	87
20) Methylene Chloride	2.794	84	3218	3.299	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022485.D
Acq On : 06 Jun 2021 18:09
Operator : JC/MD
Sample : PB136837ZHE#13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136837ZHE#13

Quant Time: Jun 07 04:26:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

