

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022486.D
 Acq On : 06 Jun 2021 18:33
 Operator : JC/MD
 Sample : PB136837ZHE#14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136837ZHE#14

Quant Time: Jun 07 04:26:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	132670	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118194	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47960	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54064	53.984	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.960%
35) Dibromofluoromethane	5.397	113	38545	45.814	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.620%
50) Toluene-d8	8.653	98	162211	50.051	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.100%
62) 4-Bromofluorobenzene	11.085	95	57814	48.263	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.520%
Target Compounds						
16) Acetone	2.386	43	3647	8.363	ug/l #	85
20) Methylene Chloride	2.794	84	2667	2.755	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022486.D
Acq On : 06 Jun 2021 18:33
Operator : JC/MD
Sample : PB136837ZHE#14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136837ZHE#14

Quant Time: Jun 07 04:26:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

