

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022487.D
 Acq On : 06 Jun 2021 18:56
 Operator : JC/MD
 Sample : M2598-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 23

Manual Integrations
 APPROVED

MMDadoda
 6/7/2021 4:54:17 PM

Quant Time: Jun 07 04:26:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	63251	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	128734	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42545	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52033	53.329	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	106.660%		
35) Dibromofluoromethane	5.391	113	37973	46.514	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.020%		
50) Toluene-d8	8.653	98	154726	49.201	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.400%		
62) 4-Bromofluorobenzene	11.085	95	52809	45.433	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.860%		
Target Compounds					Qvalue	
16) Acetone	2.386	43	2734	6.435	ug/l	89
20) Methylene Chloride	2.794	84	11076	11.742	ug/l	90
30) Chloroform	5.092	83	2054m	1.321	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022487.D
Acq On : 06 Jun 2021 18:56
Operator : JC/MD
Sample : M2598-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
23

Manual Integrations
APPROVED

MMDadoda
6/7/2021 4:54:17 PM

Quant Time: Jun 07 04:26:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

