

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022489.D
 Acq On : 06 Jun 2021 19:43
 Operator : JC/MD
 Sample : M2598-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 13

Quant Time: Jun 07 04:27:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59753	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.775	114	123247	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	109483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43929	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.965	65	49684	53.903	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.800%
35) Dibromofluoromethane	5.398	113	36551	46.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.540%
50) Toluene-d8	8.653	98	148819	49.430	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.860%
62) 4-Bromofluorobenzene	11.086	95	53771	48.320	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.640%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	10049	25.038	ug/l	93
20) Methylene Chloride	2.794	84	14463	16.230	ug/l	90
30) Chloroform	5.105	83	1482	1.009	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022489.D
Acq On : 06 Jun 2021 19:43
Operator : JC/MD
Sample : M2598-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
13

Quant Time: Jun 07 04:27:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

