

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022490.D
 Acq On : 06 Jun 2021 20:07
 Operator : JC/MD
 Sample : M2598-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 17

Quant Time: Jun 07 04:27:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60362	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	124026	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49946	53.640	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.280%			
35) Dibromofluoromethane	5.397	113	36771	46.752	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.500%			
50) Toluene-d8	8.653	98	152682	50.394	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.780%			
62) 4-Bromofluorobenzene	11.085	95	53093	47.411	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.820%			
Target Compounds						Qvalue
20) Methylene Chloride	2.794	84	14916	16.570	ug/l	# 87
30) Chloroform	5.105	83	1760	1.186	ug/l	# 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022490.D
Acq On : 06 Jun 2021 20:07
Operator : JC/MD
Sample : M2598-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
17

Quant Time: Jun 07 04:27:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

