

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022491.D
 Acq On : 06 Jun 2021 20:30
 Operator : JC/MD
 Sample : PB136829TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136829TB

Quant Time: Jun 07 04:27:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	137083	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47898	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55849	55.137	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.280%
35) Dibromofluoromethane	5.397	113	39809	45.794	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.580%
50) Toluene-d8	8.653	98	170014	50.770	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.540%
62) 4-Bromofluorobenzene	11.085	95	59883	48.381	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.760%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6233	14.132	ug/l	91
20) Methylene Chloride	2.794	84	3828	3.909	ug/l	88
43) Isopropyl Acetate	6.355	43	13939	5.484	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022491.D
Acq On : 06 Jun 2021 20:30
Operator : JC/MD
Sample : PB136829TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136829TB

Quant Time: Jun 07 04:27:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

