

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022492.D
 Acq On : 06 Jun 2021 20:53
 Operator : JC/MD
 Sample : PB136829ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136829ZHE#01

Quant Time: Jun 07 04:28:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60001	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	123583	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41912	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50487	54.548	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.100%
35) Dibromofluoromethane	5.391	113	36827	46.991	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.980%
50) Toluene-d8	8.653	98	153570	50.869	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.740%
62) 4-Bromofluorobenzene	11.085	95	53183	47.661	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.320%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6094	15.121	ug/l	98
20) Methylene Chloride	2.794	84	4243	4.742	ug/l	# 87
43) Isopropyl Acetate	6.348	43	13708	5.983	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022492.D
Acq On : 06 Jun 2021 20:53
Operator : JC/MD
Sample : PB136829ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136829ZHE#01

Quant Time: Jun 07 04:28:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

