

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022493.D
 Acq On : 06 Jun 2021 21:17
 Operator : JC/MD
 Sample : PB136829ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136829ZHE#02

Quant Time: Jun 07 04:28:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61023	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	128090	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43827	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52886	56.183	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.360%
35) Dibromofluoromethane	5.397	113	38211	47.041	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.080%
50) Toluene-d8	8.653	98	158784	50.746	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.500%
62) 4-Bromofluorobenzene	11.085	95	56015	48.433	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.860%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6010	14.663	ug/l	# 84
20) Methylene Chloride	2.788	84	4096	4.501	ug/l	95
43) Isopropyl Acetate	6.348	43	14123	5.947	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022493.D
Acq On : 06 Jun 2021 21:17
Operator : JC/MD
Sample : PB136829ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136829ZHE#02

Quant Time: Jun 07 04:28:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

