

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022495.D
 Acq On : 06 Jun 2021 22:04
 Operator : JC/MD
 Sample : PB136829ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136829ZHE#04

Quant Time: Jun 07 04:28:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60160	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	124799	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	109347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38481	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	51105	55.069	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.140%
35) Dibromofluoromethane	5.397	113	37070	46.840	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.680%
50) Toluene-d8	8.653	98	153823	50.457	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.920%
62) 4-Bromofluorobenzene	11.085	95	50732	45.022	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.040%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	6371	15.767	ug/l	96
18) Methyl Acetate	2.715	43	1515	1.410	ug/l #	86
20) Methylene Chloride	2.794	84	4161	4.638	ug/l #	86
43) Isopropyl Acetate	6.354	43	14222	6.146	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022495.D
Acq On : 06 Jun 2021 22:04
Operator : JC/MD
Sample : PB136829ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136829ZHE#04

Quant Time: Jun 07 04:28:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

