

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022497.D
 Acq On : 06 Jun 2021 22:51
 Operator : JC/MD
 Sample : PB136829ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136829ZHE#06

Quant Time: Jun 07 04:29:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	124648	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44415	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50759	54.988	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 109.980%			
35) Dibromofluoromethane	5.397	113	37301	47.189	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.380%			
50) Toluene-d8	8.653	98	154616	50.778	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.560%			
62) 4-Bromofluorobenzene	11.085	95	54631	48.541	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.080%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	6236	15.515	ug/l	99
18) Methyl Acetate	2.721	43	1602	1.499	ug/l	93
20) Methylene Chloride	2.794	84	5277	5.913	ug/l	94
43) Isopropyl Acetate	6.348	43	13919	6.023	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022497.D
Acq On : 06 Jun 2021 22:51
Operator : JC/MD
Sample : PB136829ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136829ZHE#06

Quant Time: Jun 07 04:29:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

