

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022498.D
 Acq On : 06 Jun 2021 23:14
 Operator : JC/MD
 Sample : PB136829ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136829ZHE#07

Quant Time: Jun 07 04:29:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61975	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	127803	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47824	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51836	54.221	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 108.440%			
35) Dibromofluoromethane	5.397	113	37726	46.549	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.100%			
50) Toluene-d8	8.653	98	157792	50.542	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.080%			
62) 4-Bromofluorobenzene	11.085	95	57029	49.421	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.840%			
Target Compounds						
16) Acetone	2.386	43	5845	14.041	ug/l	Qvalue 96
18) Methyl Acetate	2.709	43	1917	1.732	ug/l	# 85
20) Methylene Chloride	2.794	84	5335	5.772	ug/l	95
43) Isopropyl Acetate	6.348	43	14018	5.916	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022498.D
Acq On : 06 Jun 2021 23:14
Operator : JC/MD
Sample : PB136829ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136829ZHE#07

Quant Time: Jun 07 04:29:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

