

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022503.D
 Acq On : 07 Jun 2021 01:58
 Operator : JC/MD
 Sample : VX0606WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0606WBL02

Quant Time: Jun 07 04:31:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60715	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	125666	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	108651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41690	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	50608	54.035	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.080%	
35) Dibromofluoromethane	5.397	113	36098	45.297	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 90.600%	
50) Toluene-d8	8.653	98	150911	49.160	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.320%	
62) 4-Bromofluorobenzene	11.085	95	51773	45.629	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022503.D
Acq On : 07 Jun 2021 01:58
Operator : JC/MD
Sample : VX0606WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0606WBL02

Quant Time: Jun 07 04:31:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

