

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022506.D
 Acq On : 07 Jun 2021 03:09
 Operator : JC/MD
 Sample : PB136857TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136857TB

Quant Time: Jun 07 06:17:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61480	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	123584	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43641	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49960	52.680	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.360%
35) Dibromofluoromethane	5.397	113	36696	46.823	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.640%
50) Toluene-d8	8.653	98	155055	51.361	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.720%
62) 4-Bromofluorobenzene	11.085	95	55711	49.927	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.860%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	8082	19.571	ug/l	88
20) Methylene Chloride	2.794	84	4132	4.507	ug/l #	88
43) Isopropyl Acetate	6.354	43	11061	4.827	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022506.D
Acq On : 07 Jun 2021 03:09
Operator : JC/MD
Sample : PB136857TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136857TB

Quant Time: Jun 07 06:17:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

