

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022510.D
 Acq On : 07 Jun 2021 04:43
 Operator : JC/MD
 Sample : PB136857ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136857ZHE#04

Quant Time: Jun 07 06:18:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60193	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.775	114	125931	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	115915	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51317	55.268	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	110.540%	
35) Dibromofluoromethane	5.397	113	37567	47.041	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.080%	
50) Toluene-d8	8.653	98	158764	51.609	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	103.220%	
62) 4-Bromofluorobenzene	11.085	95	56529	49.715	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	99.440%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6502	16.082	ug/l	91
18) Methyl Acetate	2.715	43	1609	1.496	ug/l #	71
20) Methylene Chloride	2.800	84	5452	6.073	ug/l	95
43) Isopropyl Acetate	6.348	43	13063	5.595	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022510.D
Acq On : 07 Jun 2021 04:43
Operator : JC/MD
Sample : PB136857ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136857ZHE#04

Quant Time: Jun 07 06:18:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

