

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022513.D
 Acq On : 07 Jun 2021 05:53
 Operator : JC/MD
 Sample : PB136857ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136857ZHE#07

Quant Time: Jun 07 06:19:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61666	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	127129	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117895	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51452	54.089	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.180%
35) Dibromofluoromethane	5.397	113	36855	45.715	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.420%
50) Toluene-d8	8.653	98	158635	51.081	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.160%
62) 4-Bromofluorobenzene	11.085	95	57230	49.858	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.720%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	7794	18.817	ug/l	95
18) Methyl Acetate	2.715	43	2082	1.890	ug/l #	87
20) Methylene Chloride	2.788	84	4269	4.642	ug/l	91
43) Isopropyl Acetate	6.348	43	11359	4.819	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022513.D
Acq On : 07 Jun 2021 05:53
Operator : JC/MD
Sample : PB136857ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136857ZHE#07

Quant Time: Jun 07 06:19:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

