

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022514.D
 Acq On : 07 Jun 2021 06:17
 Operator : JC/MD
 Sample : PB136857ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136857ZHE#08

Quant Time: Jun 07 07:03:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	57615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	122559	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	110117	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50679	57.023	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.040%
35) Dibromofluoromethane	5.397	113	36406	46.842	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.680%
50) Toluene-d8	8.653	98	150761	50.356	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.720%
62) 4-Bromofluorobenzene	11.085	95	52378	47.332	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.660%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	8908	23.019	ug/l #	87
18) Methyl Acetate	2.715	43	1565	1.521	ug/l #	81
20) Methylene Chloride	2.794	84	4476	5.209	ug/l	93
43) Isopropyl Acetate	6.354	43	12184	5.362	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022514.D
Acq On : 07 Jun 2021 06:17
Operator : JC/MD
Sample : PB136857ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136857ZHE#08

Quant Time: Jun 07 07:03:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

