

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022516.D
 Acq On : 07 Jun 2021 07:04
 Operator : JC/MD
 Sample : PB136857ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136857ZHE#10

Quant Time: Jun 07 08:35:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	57197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	118828	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	107349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47891	54.280	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 108.560%			
35) Dibromofluoromethane	5.397	113	34596	45.911	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.820%			
50) Toluene-d8	8.653	98	145723	50.201	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.400%			
62) 4-Bromofluorobenzene	11.085	95	50836	47.381	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.760%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	8530	22.203	ug/l	90
18) Methyl Acetate	2.709	43	1603	1.569	ug/l #	84
20) Methylene Chloride	2.788	84	4574	5.362	ug/l	87
43) Isopropyl Acetate	6.348	43	12233	5.552	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022516.D
Acq On : 07 Jun 2021 07:04
Operator : JC/MD
Sample : PB136857ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136857ZHE#10

Quant Time: Jun 07 08:35:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

