

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022519.D
 Acq On : 07 Jun 2021 08:14
 Operator : JC/MD
 Sample : PB136857ZHE#13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136857ZHE#13

Quant Time: Jun 07 08:36:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	61290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	125139	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51837	54.828	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.660%	
35) Dibromofluoromethane	5.391	113	36668	46.206	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 92.420%	
50) Toluene-d8	8.653	98	156203	51.098	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.200%	
62) 4-Bromofluorobenzene	11.085	95	54885	48.575	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.160%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	8773	21.311	ug/l	93
18) Methyl Acetate	2.715	43	1467	1.340	ug/l	94
20) Methylene Chloride	2.794	84	4437	4.854	ug/l	93
43) Isopropyl Acetate	6.348	43	12150	5.237	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022519.D
Acq On : 07 Jun 2021 08:14
Operator : JC/MD
Sample : PB136857ZHE#13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136857ZHE#13

Quant Time: Jun 07 08:36:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

