

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022522.D
 Acq On : 07 Jun 2021 09:25
 Operator : JC/MD
 Sample : PB136857ZHE#16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136857ZHE#16

Quant Time: Jun 07 10:08:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59574	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	122487	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	109758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39361	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50812	55.292	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.580%
35) Dibromofluoromethane	5.397	113	36217	46.626	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.260%
50) Toluene-d8	8.653	98	151114	50.504	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.000%
62) 4-Bromofluorobenzene	11.085	95	51353	46.433	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.860%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	8519	21.290	ug/l	95
20) Methylene Chloride	2.794	84	4332	4.876	ug/l	# 78
25) 2-Butanone	4.574	43	1530	2.363	ug/l	99
43) Isopropyl Acetate	6.348	43	12573	5.536	ug/l	92
80) 1,3,5-Trimethylbenzene	11.451	105	2299	0.833	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
Data File : VX022522.D
Acq On : 07 Jun 2021 09:25
Operator : JC/MD
Sample : PB136857ZHE#16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136857ZHE#16

Quant Time: Jun 07 10:08:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

