



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060621\
 Data File : VX022477.D
 Acq On : 06 Jun 2021 14:28
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Title : SW846 8260
 Last Update : Wed May 19 14:23:10 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	6120	PASS
75	95	30	60	51.3	16251	PASS
95	95	100	100	100.0	31651	PASS
96	95	5	9	6.8	2165	PASS
173	174	0.00	2	0.9	176	PASS
174	95	50	100	62.9	19899	PASS
175	174	5	9	7.3	1454	PASS
176	174	95	101	96.5	19208	PASS
177	176	5	9	6.4	1231	PASS