

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
 Data File : VX029237.D
 Acq On : 06 Jun 2022 14:24
 Operator : JC/MD
 Sample : N3174-03DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 INF0622DL

Quant Time: Jun 07 06:04:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	296776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	546259	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	489408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	218791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	195860	58.526	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.060%
35) Dibromofluoromethane	5.391	113	171826	47.356	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.720%
50) Toluene-d8	8.653	98	648415	48.861	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.720%
62) 4-Bromofluorobenzene	11.085	95	234172	45.650	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.300%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.470	56	31832	7.316	ug/l	95
39) Methylcyclohexane	7.379	83	16629	2.881	ug/l	89
40) Benzene	6.043	78	245593	17.761	ug/l	97
52) Toluene	8.720	92	15041	1.628	ug/l	93
67) Ethyl Benzene	10.195	91	8856	0.524	ug/l	95
68) m/p-Xylenes	10.305	106	32217	4.749	ug/l	99
69) o-Xylene	10.640	106	92545	13.803	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	128243	10.188	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	9671	0.770	ug/l	98
95) Naphthalene	13.780	128	153423	10.568	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
Data File : VX029237.D
Acq On : 06 Jun 2022 14:24
Operator : JC/MD
Sample : N3174-03DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
INF0622DL

Quant Time: Jun 07 06:04:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

