

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
 Data File : VX029238.D
 Acq On : 06 Jun 2022 14:47
 Operator : JC/MD
 Sample : N3143-07RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20220408-FIELD-BLANK-1RE

Quant Time: Jun 07 06:04:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	318299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	595636	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	591062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	274685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	216177	60.229	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	120.460%
35) Dibromofluoromethane	5.391	113	195291	49.361	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.720%
50) Toluene-d8	8.653	98	708967	48.996	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.000%
62) 4-Bromofluorobenzene	11.079	95	286675	51.252	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.500%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	4754	3.364	ug/l	98
20) Methylene Chloride	2.788	84	1096	0.333	ug/l #	84
68) m/p-Xylenes	10.311	106	2707	0.330	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
Data File : VX029238.D
Acq On : 06 Jun 2022 14:47
Operator : JC/MD
Sample : N3143-07RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20220408-FIELD-BLANK-1RE

Quant Time: Jun 07 06:04:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

