

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
 Data File : VX029241.D
 Acq On : 06 Jun 2022 15:57
 Operator : JC/MD
 Sample : N3195-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 07 06:05:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	304078	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	575411	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	561450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	257067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	210154	61.289	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	122.580%
35) Dibromofluoromethane	5.391	113	190410	49.819	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.640%
50) Toluene-d8	8.653	98	678719	48.554	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.100%
62) 4-Bromofluorobenzene	11.079	95	272132	50.362	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.720%
Target Compounds						
						Qvalue
3) Chloromethane	1.288	50	822	0.297	ug/l	# 86
16) Acetone	2.386	43	2850	2.111	ug/l	91
20) Methylene Chloride	2.788	84	1042	0.331	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
Data File : VX029241.D
Acq On : 06 Jun 2022 15:57
Operator : JC/MD
Sample : N3195-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 07 06:05:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

