

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
 Data File : VX029246.D
 Acq On : 06 Jun 2022 18:17
 Operator : JC/MD
 Sample : N3040-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-TB-20220524RE

Quant Time: Jun 07 06:06:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	304360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	572370	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	563549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	262999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	210631	61.371	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 122.740%	
35) Dibromofluoromethane	5.385	113	191104	50.267	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.540%	
50) Toluene-d8	8.653	98	681541	49.015	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.020%	
62) 4-Bromofluorobenzene	11.085	95	275364	51.231	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.460%	
Target Compounds						
						Qvalue
3) Chloromethane	1.295	50	941	0.340	ug/l	97
16) Acetone	2.386	43	2724	2.016	ug/l	92
20) Methylene Chloride	2.794	84	1404	0.446	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
Data File : VX029246.D
Acq On : 06 Jun 2022 18:17
Operator : JC/MD
Sample : N3040-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-TB-20220524RE

Quant Time: Jun 07 06:06:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

