

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
 Data File : VX029250.D
 Acq On : 06 Jun 2022 19:51
 Operator : JC/MD
 Sample : N3189-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20220428-SPLP-AMENDED-COMPOSITE-H

Quant Time: Jun 07 06:06:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	295593	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	566566	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	550471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	258595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	207731	62.321	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	124.640%
35) Dibromofluoromethane	5.391	113	119859	31.850	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	63.700%#
50) Toluene-d8	8.653	98	673346	48.922	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.840%
62) 4-Bromofluorobenzene	11.085	95	271904	51.106	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.220%
Target Compounds						
16) Acetone	2.380	43	46630	35.526	ug/l	95
20) Methylene Chloride	2.794	84	6691	2.186	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
Data File : VX029250.D
Acq On : 06 Jun 2022 19:51
Operator : JC/MD
Sample : N3189-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20220428-SPLP-AMENDED-COMPOSITE-H

Quant Time: Jun 07 06:06:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

