

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
 Data File : VX029260.D
 Acq On : 07 Jun 2022 00:53
 Operator : JC/MD
 Sample : N3166-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB07-20220601

Quant Time: Jun 07 06:09:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	312041	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	592024	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	583687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	267413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	217966	61.945	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 123.880%			
35) Dibromofluoromethane	5.391	113	196313	49.922	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.840%			
50) Toluene-d8	8.653	98	708544	49.265	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 98.540%			
62) 4-Bromofluorobenzene	11.079	95	284557	51.184	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.360%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	25763	18.594	ug/l	97
20) Methylene Chloride	2.794	84	2699	0.835	ug/l #	92
21) trans-1,2-Dichloroethene	3.099	96	1337	0.421	ug/l	92
27) cis-1,2-Dichloroethene	4.489	96	1916	0.506	ug/l	89
44) Trichloroethene	7.129	130	7057	0.899	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
Data File : VX029260.D
Acq On : 07 Jun 2022 00:53
Operator : JC/MD
Sample : N3166-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB07-20220601

Quant Time: Jun 07 06:09:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

