

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
 Data File : VX029262.D
 Acq On : 07 Jun 2022 01:39
 Operator : JC/MD
 Sample : N3166-28
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20220531

Quant Time: Jun 07 06:09:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	306551	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	580572	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	566456	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	259305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	215311	62.286	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 124.580%	
35) Dibromofluoromethane	5.391	113	194143	50.344	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.680%	
50) Toluene-d8	8.653	98	690471	48.955	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.920%	
62) 4-Bromofluorobenzene	11.085	95	276256	50.671	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.340%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	19736	14.499	ug/l	100
20) Methylene Chloride	2.788	84	3494	1.101	ug/l	97
25) 2-Butanone	4.574	43	3735	1.701	ug/l	# 88
27) cis-1,2-Dichloroethene	4.495	96	2970	0.799	ug/l	86
44) Trichloroethene	7.129	130	4367	0.304	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
Data File : VX029262.D
Acq On : 07 Jun 2022 01:39
Operator : JC/MD
Sample : N3166-28
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20220531

Quant Time: Jun 07 06:09:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

