

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
 Data File : VX029252.D
 Acq On : 06 Jun 2022 21:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 07 06:06:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.435	0.473	-8.7	93	0.00
3 P	Chloromethane	0.455	0.491	-7.9	99	0.00
4 C	Vinyl Chloride	0.500	0.548	-9.6#	95	0.00
5 T	Bromomethane	0.254	0.240	5.5	88	0.00
6 T	Chloroethane	0.335	0.347	-3.6	91	0.00
7 T	Trichlorofluoromethane	0.777	0.826	-6.3	94	0.00
8 T	Diethyl Ether	0.298	0.322	-8.1	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.461	0.505	-9.5	99	0.00
10 T	Methyl Iodide	0.505	0.603	-19.4	93	0.00
11 T	Tert butyl alcohol	0.131	0.126	3.8	90	0.00
12 CM	1,1-Dichloroethene	0.451	0.487	-8.0#	97	0.00
13 T	Acrolein	0.059	0.052	11.9	77	0.00
14 T	Allyl chloride	0.677	0.790	-16.7	104	0.00
15 T	Acrylonitrile	0.269	0.323	-20.1	106	0.00
16 T	Acetone	0.222	0.264	-18.9	108	0.00
17 T	Carbon Disulfide	1.236	1.119	9.5	83	0.00
18 T	Methyl Acetate	0.574	0.720	-25.4#	114	0.00
19 T	Methyl tert-butyl Ether	1.529	1.787	-16.9	102	0.00
20 T	Methylene Chloride	0.518	0.577	-11.4	101	0.00
21 T	trans-1,2-Dichloroethene	0.508	0.551	-8.5	97	0.00
22 T	Diisopropyl ether	1.354	1.690	-24.8	109	0.00
23 T	Vinyl Acetate	1.131	1.433	-26.7#	107	0.00
24 P	1,1-Dichloroethane	0.838	0.986	-17.7	104	0.00
25 T	2-Butanone	0.358	0.441	-23.2	110	0.00
26 T	2,2-Dichloropropane	0.687	0.626	8.9	80	0.00
27 T	cis-1,2-Dichloroethene	0.606	0.659	-8.7	99	0.00
28 T	Bromochloromethane	0.337	0.410	-21.7	115	0.00
29 T	Tetrahydrofuran	0.230	0.291	-26.5#	110	0.00
30 C	Chloroform	0.928	1.072	-15.5#	101	0.00
31 T	Cyclohexane	0.733	0.879	-19.9	105	0.00
32 T	1,1,1-Trichloroethane	0.841	0.934	-11.1	97	0.00
33 S	1,2-Dichloroethane-d4	0.564	0.615	-9.0	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.332	0.306	7.8	90	0.00
36 T	1,1-Dichloropropene	0.427	0.447	-4.7	102	0.00
37 T	Ethyl Acetate	0.414	0.478	-15.5	107	0.00
38 T	Carbon Tetrachloride	0.488	0.465	4.7	91	0.00
39 T	Methylcyclohexane	0.528	0.543	-2.8	98	0.00
40 TM	Benzene	1.266	1.348	-6.5	102	0.00
41 T	Methacrylonitrile	0.224	0.268	-19.6	110	0.00
42 TM	1,2-Dichloroethane	0.429	0.468	-9.1	104	0.00
43 T	Isopropyl Acetate	0.648	0.722	-11.4	105	0.00
44 TM	Trichloroethene	0.434	0.365	15.9	93	0.00
45 C	1,2-Dichloropropane	0.305	0.334	-9.5#	104	0.00
46 T	Dibromomethane	0.238	0.248	-4.2	101	0.00
47 T	Bromodichloromethane	0.468	0.474	-1.3	95	0.00
48 T	Methyl methacrylate	0.315	0.365	-15.9	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
 Data File : VX029252.D
 Acq On : 06 Jun 2022 21:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 07 06:06:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	96	0.02
50 S	Toluene-d8	1.215	1.142	6.0	92	0.00
51 T	4-Methyl-2-Pentanone	0.431	0.488	-13.2	107	0.00
52 CM	Toluene	0.845	0.884	-4.6#	100	0.00
53 T	t-1,3-Dichloropropene	0.481	0.470	2.3	90	0.00
54 T	cis-1,3-Dichloropropene	0.527	0.532	-0.9	95	0.00
55 T	1,1,2-Trichloroethane	0.352	0.364	-3.4	99	0.00
56 T	Ethyl methacrylate	0.499	0.550	-10.2	102	0.00
57 T	1,3-Dichloropropane	0.561	0.599	-6.8	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.237	0.267	-12.7	102	0.00
59 T	2-Hexanone	0.335	0.374	-11.6	104	0.00
60 T	Dibromochloromethane	0.400	0.365	8.8	86	0.00
61 T	1,2-Dibromoethane	0.380	0.381	-0.3	97	0.00
62 S	4-Bromofluorobenzene	0.470	0.435	7.4	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.376	0.349	7.2	89	0.00
65 PM	Chlorobenzene	1.027	1.012	1.5	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.396	0.376	5.1	91	0.00
67 C	Ethyl Benzene	1.726	1.830	-6.0#	98	0.00
68 T	m/p-Xylenes	0.693	0.708	-2.2	94	0.00
69 T	o-Xylene	0.685	0.694	-1.3	93	0.00
70 T	Styrene	1.140	1.160	-1.8	92	0.00
71 P	Bromoform	0.342	0.274	19.9	75	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.360	3.756	-11.8	95	0.00
74 T	N-amyl acetate	1.088	1.319	-21.2	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.133	1.244	-9.8	95	0.00
76 T	1,2,3-Trichloropropane	1.014	1.109	-9.4	95	0.00
77 T	Bromobenzene	0.896	0.871	2.8	86	0.00
78 T	n-propylbenzene	3.737	4.264	-14.1	96	0.00
79 T	2-Chlorotoluene	2.362	2.613	-10.6	96	0.00
80 T	1,3,5-Trimethylbenzene	2.877	3.181	-10.6	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.345	0.320	7.2	78	0.00
82 T	4-Chlorotoluene	2.704	2.992	-10.7	95	0.00
83 T	tert-Butylbenzene	2.940	3.105	-5.6	90	0.00
84 T	1,2,4-Trimethylbenzene	2.871	3.198	-11.4	94	0.00
85 T	sec-Butylbenzene	3.452	3.875	-12.3	94	0.00
86 T	p-Isopropyltoluene	3.018	3.245	-7.5	90	0.00
87 T	1,3-Dichlorobenzene	1.689	1.653	2.1	85	0.00
88 T	1,4-Dichlorobenzene	1.740	1.675	3.7	85	0.00
89 T	n-Butylbenzene	2.449	2.751	-12.3	94	0.00
90 T	Hexachloroethane	0.549	0.515	6.2	80	0.00
91 T	1,2-Dichlorobenzene	1.664	1.638	1.6	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.253	0.272	-7.5	90	0.00
93 T	1,2,4-Trichlorobenzene	1.038	1.008	2.9	85	0.00
94 T	Hexachlorobutadiene	0.475	0.400	15.8	77	0.00
95 T	Naphthalene	3.318	3.777	-13.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060622\
Data File : VX029252.D
Acq On : 06 Jun 2022 21:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 07 06:06:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.025	1.026	-0.1	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6