

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060624\
 Data File : VX041769.D
 Acq On : 06 Jun 2024 12:00
 Operator : JC/MD
 Sample : P2749-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-10D

Quant Time: Jun 07 00:58:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	154730	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	236175	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229091	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111802	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91325	48.124	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.240%
35) Dibromofluoromethane	5.385	113	75132	46.011	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.020%
50) Toluene-d8	8.647	98	277153	48.576	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	102500	48.675	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.340%
Target Compounds						
						Qvalue
30) Chloroform	5.093	83	27192	9.877	ug/l	99
44) Trichloroethene	7.135	130	812	0.483	ug/l	76
64) Tetrachloroethene	9.275	164	707	0.412	ug/l	# 90
65) Chlorobenzene	10.079	112	4862	0.989	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060624\
Data File : VX041769.D
Acq On : 06 Jun 2024 12:00
Operator : JC/MD
Sample : P2749-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-10D

Quant Time: Jun 07 00:58:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

