

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060624\
 Data File : VX041772.D
 Acq On : 06 Jun 2024 13:10
 Operator : JC/MD
 Sample : P2749-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-12I

Quant Time: Jun 07 01:00:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	126531	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	196752	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76319	49.179	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.360%
35) Dibromofluoromethane	5.385	113	62369	45.848	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.700%
50) Toluene-d8	8.647	98	229907	48.369	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.740%
62) 4-Bromofluorobenzene	11.079	95	85529	48.754	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.500%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.002	59	2486	9.056	ug/l	# 91
16) Acetone	2.392	43	3694	2.732	ug/l	# 89
24) 1,1-Dichloroethane	3.611	63	11075	5.165	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	7873	5.628	ug/l	82
29) Tetrahydrofuran	5.038	42	3523	5.792	ug/l	95
40) Benzene	6.032	78	3844	0.747	ug/l	97
44) Trichloroethene	7.135	130	1428	1.020	ug/l	96
64) Tetrachloroethene	9.275	164	1173	0.826	ug/l	# 81
65) Chlorobenzene	10.080	112	122655	30.170	ug/l	100
79) 2-Chlorotoluene	11.366	91	4297	1.065	ug/l	95
87) 1,3-Dichlorobenzene	11.969	146	11244	3.732	ug/l	95
88) 1,4-Dichlorobenzene	12.043	146	44002	14.168	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	20844	6.792	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060624\
Data File : VX041772.D
Acq On : 06 Jun 2024 13:10
Operator : JC/MD
Sample : P2749-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-12I

Quant Time: Jun 07 01:00:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

