

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX060624\
 Data File : VX041763.D
 Acq On : 06 Jun 2024 09:32
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 07 00:55:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	107	0.00
2 T	Dichlorodifluoromethane	50.000	45.806	8.4	96	0.00
3 P	Chloromethane	50.000	43.343	13.3	98	0.00
4 C	Vinyl Chloride	50.000	44.049	11.9#	97	0.00
5 T	Bromomethane	50.000	43.496	13.0	94	0.00
6 T	Chloroethane	50.000	41.205	17.6	100	0.00
7 T	Trichlorofluoromethane	50.000	44.435	11.1	99	0.00
8 T	Diethyl Ether	50.000	43.988	12.0	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	46.092	7.8	101	0.00
10 T	Methyl Iodide	50.000	45.225	9.5	96	0.00
11 T	Tert butyl alcohol	250.000	244.609	2.2	111	0.00
12 CM	1,1-Dichloroethene	50.000	44.551	10.9#	97	0.00
13 T	Acrolein	250.000	258.450	-3.4	114	0.00
14 T	Allyl chloride	50.000	47.598	4.8	99	0.00
15 T	Acrylonitrile	250.000	240.199	3.9	103	0.00
16 T	Acetone	250.000	297.375	-18.9	130	0.00
17 T	Carbon Disulfide	50.000	42.396	15.2	90	0.00
18 T	Methyl Acetate	50.000	50.484	-1.0	110	0.00
19 T	Methyl tert-butyl Ether	50.000	46.006	8.0	98	0.00
20 T	Methylene Chloride	50.000	41.235	17.5	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.343	11.3	96	0.00
22 T	Diisopropyl ether	50.000	47.013	6.0	99	0.00
23 T	Vinyl Acetate	250.000	241.371	3.5	99	0.00
24 P	1,1-Dichloroethane	50.000	45.719	8.6	98	0.00
25 T	2-Butanone	250.000	264.071	-5.6	112	0.00
26 T	2,2-Dichloropropane	50.000	46.499	7.0	101	0.00
27 T	cis-1,2-Dichloroethene	50.000	45.879	8.2	99	0.00
28 T	Bromochloromethane	50.000	42.091	15.8	93	0.00
29 T	Tetrahydrofuran	250.000	245.345	1.9	105	0.00
30 C	Chloroform	50.000	45.975	8.0#	99	0.00
31 T	Cyclohexane	50.000	45.368	9.3	99	0.00
32 T	1,1,1-Trichloroethane	50.000	46.434	7.1	99	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.008	4.0	105	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	108	0.00
35 S	Dibromofluoromethane	50.000	47.846	4.3	105	0.00
36 T	1,1-Dichloropropene	50.000	44.838	10.3	98	0.00
37 T	Ethyl Acetate	50.000	47.346	5.3	102	0.00
38 T	Carbon Tetrachloride	50.000	45.626	8.7	97	0.00
39 T	Methylcyclohexane	50.000	47.193	5.6	100	0.00
40 TM	Benzene	50.000	45.439	9.1	98	0.00
41 T	Methacrylonitrile	50.000	53.088	-6.2	109	0.00
42 TM	1,2-Dichloroethane	50.000	45.735	8.5	98	0.00
43 T	Isopropyl Acetate	50.000	46.953	6.1	101	0.00
44 TM	Trichloroethene	50.000	45.685	8.6	98	0.00
45 C	1,2-Dichloropropane	50.000	47.013	6.0#	99	0.00
46 T	Dibromomethane	50.000	46.269	7.5	98	0.00
47 T	Bromodichloromethane	50.000	48.234	3.5	99	0.00
48 T	Methyl methacrylate	50.000	50.222	-0.4	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060624\
 Data File : VX041763.D
 Acq On : 06 Jun 2024 09:32
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 07 00:55:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1049.898	-5.0	113	0.00
50 S	Toluene-d8	50.000	48.605	2.8	104	0.00
51 T	4-Methyl-2-Pentanone	250.000	251.631	-0.7	105	0.00
52 CM	Toluene	50.000	46.828	6.3#	99	0.00
53 T	t-1,3-Dichloropropene	50.000	49.331	1.3	100	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.276	3.4	98	0.00
55 T	1,1,2-Trichloroethane	50.000	47.074	5.9	101	0.00
56 T	Ethyl methacrylate	50.000	50.252	-0.5	101	0.00
57 T	1,3-Dichloropropane	50.000	47.026	5.9	101	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	246.488	1.4	100	0.00
59 T	2-Hexanone	250.000	267.006	-6.8	109	0.00
60 T	Dibromochloromethane	50.000	49.275	1.5	100	0.00
61 T	1,2-Dibromoethane	50.000	47.359	5.3	100	0.00
62 S	4-Bromofluorobenzene	50.000	51.646	-3.3	109	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	107	0.00
64 T	Tetrachloroethene	50.000	45.328	9.3	100	0.00
65 PM	Chlorobenzene	50.000	45.305	9.4	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.344	5.3	101	0.00
67 C	Ethyl Benzene	50.000	47.646	4.7#	101	0.00
68 T	m/p-Xylenes	100.000	94.453	5.5	99	0.00
69 T	o-Xylene	50.000	47.020	6.0	100	0.00
70 T	Styrene	50.000	49.106	1.8	102	0.00
71 P	Bromoform	50.000	48.990	2.0	100	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	107	0.00
73 T	Isopropylbenzene	50.000	47.900	4.2	102	0.00
74 T	N-ethyl acetate	50.000	49.933	0.1	102	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	46.184	7.6	102	0.00
76 T	1,2,3-Trichloropropane	50.000	48.664	2.7	105	0.00
77 T	Bromobenzene	50.000	46.106	7.8	101	0.00
78 T	n-propylbenzene	50.000	49.404	1.2	104	0.00
79 T	2-Chlorotoluene	50.000	47.019	6.0	102	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.151	3.7	102	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.194	3.6	100	0.00
82 T	4-Chlorotoluene	50.000	47.774	4.5	102	0.00
83 T	tert-Butylbenzene	50.000	47.749	4.5	102	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.933	2.1	102	0.00
85 T	sec-Butylbenzene	50.000	49.514	1.0	103	0.00
86 T	p-Isopropyltoluene	50.000	50.129	-0.3	104	0.00
87 T	1,3-Dichlorobenzene	50.000	46.996	6.0	102	0.00
88 T	1,4-Dichlorobenzene	50.000	45.808	8.4	101	0.00
89 T	n-Butylbenzene	50.000	52.342	-4.7	106	0.00
90 T	Hexachloroethane	50.000	48.194	3.6	103	0.00
91 T	1,2-Dichlorobenzene	50.000	46.379	7.2	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.523	-3.0	105	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.908	2.2	102	0.00
94 T	Hexachlorobutadiene	50.000	47.725	4.5	104	0.00
95 T	Naphthalene	50.000	50.194	-0.4	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060624\
Data File : VX041763.D
Acq On : 06 Jun 2024 09:32
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 07 00:55:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.038	1.9	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6