



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060624\
Data File : VX041762.D
Acq On : 06 Jun 2024 08:57
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Title : SW846 8260
Last Update : Thu May 30 07:36:24 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	14435	PASS
75	95	30	60	51.1	38077	PASS
95	95	100	100	100.0	74512	PASS
96	95	5	9	6.7	5005	PASS
173	174	0.00	2	0.9	613	PASS
174	95	50	100	91.8	68427	PASS
175	174	5	9	7.7	5253	PASS
176	174	95	101	97.9	67008	PASS
177	176	5	9	6.9	4618	PASS