

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060719\
 Data File : VX010128.D
 Acq On : 07 Jun 2019 22:54
 Operator : JC/SP
 Sample : K3221-07 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW309I-20190605

Quant Time: Jun 08 04:13:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	300237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	453833	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	411339	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	197767	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	135919	46.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.76%	
35) Dibromofluoromethane	5.50	113	140128	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
50) Toluene-d8	8.71	98	534844	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
62) 4-Bromofluorobenzene	11.13	95	189543	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060719\
Data File : VX010128.D
Acq On : 07 Jun 2019 22:54
Operator : JC/SP
Sample : K3221-07 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW309I-20190605

Quant Time: Jun 08 04:13:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

