

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060719\
 Data File : VX010131.D
 Acq On : 08 Jun 2019 00:04
 Operator : JC/SP
 Sample : K3221-09 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-DUP02-20190605

Quant Time: Jun 08 04:27:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	306646	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	463992	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	414434	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193983	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	139509	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
35) Dibromofluoromethane	5.50	113	141689	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
50) Toluene-d8	8.71	98	544753	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
62) 4-Bromofluorobenzene	11.13	95	192081	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060719\
Data File : VX010131.D
Acq On : 08 Jun 2019 00:04
Operator : JC/SP
Sample : K3221-09 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-DUP02-20190605

Quant Time: Jun 08 04:27:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

